

Declaration of Conformity

In accordance with UK Government guidance

This declaration is issued under the sole responsibility of the manufacturer,

Heidolph Scientific Products GmbH
Walpersdorfer Straße 12
91126 Schwabach / Germany

Product: Laboratory stirring device

Model:

Hei TORQUE Expert 100	501-61011-xx
Hei TORQUE Expert 200	501-62011-xx
Hei TORQUE Expert 400	501-64011-xx
Hei TORQUE Ultimate 100	501-61021-xx
Hei TORQUE Ultimate 200	501-62021-xx
Hei TORQUE Ultimate 400	501-64021-xx
Hei TORQUE Core	501-60411-xx

Description:

Overhead stirrer to work with stirring tools eg. impeller.

The object of the declaration described above is in conformity with the relevant UK Statutory Instruments (and their amendments):

2008 No. 1597

The Supply of Machinery (Safety) Regulations

2008 2016 No. 1091

The Electromagnetic Compatibility Regulations 2016

2012 No. 3032

The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012

and complies with the following technical standards :

EN ISO 12100:2010, EN 61326-1:2013, EN 61010-1:2010, EN IEC 63000:2018

UK Authorised Representative (for authorities only):

ProductIP (UK) Ltd.
8. Northumberland Av.
London WC2N 5BY

Signed for and on behalf of Heidolph Scientific Products GmbH
Walpersdorfer Straße 12, 91126 Schwabach / Germany

Schwabach, 13.09.2024



Andreas Hahn
Managing Director



Jörg Ziel
Quality Manager